

DOPE DOG (2)

Sample ID: BIA251229S0656
Strain: hLOT#091725
Harvest Lot:
Matrix: Plant
Type: Flower - Cured
Sample Size: 2 g
Lot#:

Produced:
Collected:
Received: 12/29/2025
Completed: 01/09/2026
Batch#:

Client
BERN LIVING ORGANICS, LLC
Lic. # CLTV0089
PO BOX 3418
BURLINGTON, VT 05408



Summary

Test	Date Tested	Result
Sample		Complete
Moisture	01/05/2026	15.70% - Complete
Water Activity	01/05/2026	0.725 aw - Complete
Terpenes	01/06/2026	Complete




Luke Emerson-Mason
Laboratory Director
01/09/2026

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com



DOPE DOG (2)

Sample ID: BIA251229S0656
Strain: hLOT#091725
Harvest Lot:
Matrix: Plant
Type: Flower - Cured
Sample Size: 2 g
Lot#:

Produced:
Collected:
Received: 12/29/2025
Completed: 01/09/2026
Batch#:

Client:
BERN LIVING ORGANICS, LLC
Lic. # CLTV0089
PO BOX 3418
BURLINGTON, VT 05408

Terpenes

Completed

Analyte	LOQ	Results	Results
	mg/g	mg/g	%
Limonene	0.010	4.037	0.404
Ocimene	0.010	3.364	0.336
β-Caryophyllene	0.010	3.026	0.303
β-Myrcene	0.010	2.529	0.253
Linalool	0.010	2.409	0.241
β-Pinene	0.010	1.889	0.189
α-Pinene	0.010	1.443	0.144
α-Humulene	0.010	1.134	0.113
Camphene	0.010	0.273	0.027
Terpinolene	0.010	0.228	0.023
α-Bisabolol	0.010	0.054	0.005
γ-Terpinene	0.010	0.023	0.002
α-Terpinene	0.010	0.018	0.002
3-Carene	0.010	<LOQ	<LOQ
Caryophyllene Oxide	0.010	<LOQ	<LOQ
cis-Nerolidol	0.010	<LOQ	<LOQ
Eucalyptol	0.010	<LOQ	<LOQ
Geraniol	0.010	<LOQ	<LOQ
Guaiol	0.010	<LOQ	<LOQ
Isopulegol	0.010	<LOQ	<LOQ
p-Cymene	0.010	<LOQ	<LOQ
trans-Nerolidol	0.010	<LOQ	<LOQ
Total		20.428	2.043

Primary Aromas



Analyst: 048

LOQ = The lowest quantity this method can reliably detect. Any terpene that was not detected is assumed to be less than the stated LOQ (<LOQ).

Terpene Methodology: Headspace Sampler, Gas Chromatography-Mass Spectrometry (GC-MS), using Perkin Elmer Clarus® SQ8 GC MS

Reagent Blanks: < LOQs for all analytes

All results reflect dry weight of material, based on % moisture of the sample.

All moisture and water activity analysis is determined by dewpoint measurement using an AQUALAB water activity meter.




Luke Emerson-Mason
 Laboratory Director
 01/09/2026

Confident LIMS
 All Rights Reserved
coa.support@confidentlims.com
 (866) 506-5866
www.confidentlims.com

